



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

ATEQ CORPORATION
 35980 Industrial Road
 Livonia, MI 48150
 Bill Whitlow Phone: 734 838 3100

CALIBRATION

Valid To: September 30, 2012

Certificate Number: 3049.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Mechanical

Parameter/Equipment	Range	CMC ² (±)	Comments
Pressure – Gas	(0 to 1245) Pa (0 to 1) psig (0 to 3) psig (0 to 30) psig (0 to 100) psig (0 to 300) psig (0 to -14) psig	1.7 Pa 0.003 psig 0.02 psig 0.06 psig 0.15 psig 0.36 psig 0.04 psig	Gauge

¹ This laboratory offers commercial calibration service.

² Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.



The American Association for Laboratory Accreditation

World Class Accreditation

Accredited Laboratory

A2LA has accredited

A TEQ CORPORATION

Livonia, MI

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General Requirements for the Competence of Testing and Calibration Laboratories*. This laboratory also meets any additional program requirements in the field of calibration. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009).

Presented this 18th day of June 2010.





Peter Meyer

President & CEO
For the Accreditation Council
Certificate Number 3049.01
Valid to September 30, 2012

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.